

Certificate of Analysis

Description

Product Name Praxis ® Extended RAS Panel

Part Number 20012431

Lot Number

A159148

Test Conditions and Parameters

The Praxis Extended RAS Panel kit was functionally tested based on the Praxis Extended RAS Panel Assay Reference Guide (Document Number 1000000037191). The Praxis Extended RAS Panel kit passed functional quality control testing per Illumina internal requirements in addition to the results below.

Note: Flow cells are serialized and passed release testing separately via a hybridization assay.

Test Results

Metric	Specification	Result
delta Cq from No Template Control to Quality Control Template	≥ 10	Pass
Positive Control	Expected Mutations detected: KRAS Exon 2 p.Gly12Asp c.35G>A and NRAS Exon 3 p.Gln61Lys c.181C>A	Pass
Negative Control	Invalid by sequencing	Pass

Certification

This document certifies that the product(s) described above meet quality specifications.

Print Name: MARIAH AUBIN

Title: QC ANALYST 1

Signature:



Date:

06-APR-2020