

Certificate of Analysis

Description

| | | | |
|--------------|---|----------------|----------|
| Product Name | TruSeq Custom Amplicon Dx FFPE QC Box 1/1 | Catalog Number | 20006259 |
| Part Number | 20003464 | Lot Number | A122795 |

Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).


Test Results

| Metric | Specification | Result |
|----------------------------------|-------------------------|--------|
| ACD1 Concentration (ng/ μ l) | ≥ 10 | Pass |
| Slope of ACD1 standard Curve | $-3.8 \leq X \leq -3.2$ | Pass |
| R2 of ACD1 Standard Curve | ≥ 0.985 | Pass |
| NTC | ≥ 35.0 | Pass |

Certification

This document certifies that the product(s) described above meet quality specifications.

Quality Review

| | | | | | |
|------------|----------------|-----------|--|------|-------------|
| Print Name | DONNA MAE CRUZ | Signature |  | Date | 27 AUG 2018 |
|------------|----------------|-----------|--|------|-------------|