

Certificate of Analysis

Description

Product Name	TruSeq Custom Amplicon Dx FFPE QC Box 1/1	Catalog Number	20006259
Part Number	20003464	Lot Number	A125677

Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).

Test Results

Metric	Specification	Result
ACD1 Concentration (ng/μl)	≥ 10	Pass
Slope of ACD1 standard Curve	$-3.8 \leq X \leq -3.2$	Pass
R2 of ACD1 Standard Curve	≥ 0.985	Pass
NTC	≥ 35.0	Pass

Certification

This document certifies that the product(s) described above meet quality specifications.

Quality Review

Print Name	Caitlin Knudsen	Signature		Date	03-JAN-2019
------------	-----------------	-----------	--	------	-------------