

## Certificate of Analysis

### Description

Product Name	TruSeq Custom Amplicon Dx FFPE QC Box 1/1	Catalog Number	20006259
Part Number	20003464	Lot Number	A164065

### Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).

### Test Results

Metric	Specification	Result
ACD1 Concentration (ng/μl)	≥ 10	Pass
Slope of ACD1 standard Curve	$-3.8 \leq X \leq -3.2$	Pass
R2 of ACD1 Standard Curve	≥ 0.985	Pass
NTC	≥ 35.0	Pass

### Certification

This document certifies that the product(s) described above meet quality specifications.

### Quality Review

Print Name	MARIAH AUBIN	Signature	<i>Mariah Aubin</i>	<small>Electronically signed by: Mariah Aubin Reason: Author Date: Jul 9, 2021 19:41 PDT</small>	Date	01-Jul-2021
------------	--------------	-----------	---------------------	--	------	-------------