

Date : 12/09/2022

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Chad Cunningham

Re. : CU US + Canada Certificate

Type of Equipment : Robotic Pipetting Workstation
Model Designation : See Certificate
Certificate No. : CU 72228211 0001
File No. : US22XLNW 004
Engineer/Contact : Doan Phung
Standard(s) : UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) +
A1
UL 61010-2-081:2019
CSA C22.2 No. 61010-2-081:19

Dear Mr. Cunningham,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72228211 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body

Dipl.-Ing. M. Glagla
QA Certification Officer

TUV Rheinland
of North America, Inc.

295 Foster St.
Suite 100
Littleton, MA 01460

Tel 978-266-9500
Fax 978-266-9992
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72228211 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Hamilton Company
4970 Energy Way (building 4)
Reno NV 89502
USA

Test report no.: USA-DP US22XLNW 004

Client Reference: Chad Cunningham

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
UL 61010-2-081:2019
CSA C22.2 No. 61010-2-081:19

Certified Product: Robotic Pipetting Workstation

License Fee - Units

Model Designation: Infinium Amplification System#

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 600VA

Protection Class: I

Appendix: 1, 1-3

Licensed Test mark:



Date of Issue

(day/mo/yr)

09/12/2022