



Ref. Certif. No.

DE 2-035538

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Next Generation Sequencer
Name and address of the applicant	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the manufacturer	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the factory	See additional page(s)
Ratings and principal characteristics	AC 200-240V, 50/60Hz; 15A (2500W); Class I; Laser Class 1
Trademark (if any)	Illumina
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NovaSeq 6000
Additional information (if necessary may also be reported on page 2)	
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 IEC 61010-2-010:2019 IEC 61010-2-081:2019 IEC 60825-1:2014 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31790423 019

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2022-04-18

Signature:

Dipl.-Ing. M. Glagla

1. Illumina, Inc.
25861 Industrial Blvd.
Hayward CA 94545
USA
2. Venture Manufacturing Services Inc.
1621 Barber Lane
Milpitas CA 95035
USA

Additional information (if necessary)

Report Ref. No. : 31790423 019



Date: 2022-04-18

Signature:

Dipl.-Ing. M. Glagla